Evaluation of Stability of Serum on Different Storage Temperatures for Routine Chemistry Analytes

Byoung-rak An¹ and Chang-Eun Park²

¹Department of Laboratory Medicine, Yongin Severance Hospital, Yonsei University, Yongin 449-930, Korea

²Department of Biomedical Laboratory Science, Namseoul University, Cheonan 331-707, Korea

The stability of 21 routine biochemistry analytes was evaluated from the specimens which had been stored under three different temperature conditions for 30 days. Most of the 17 analytes showed significant change, however, the specimens under lower temperature showed more stability than those under higher temperature. Glucose, Albumin, γ -glutamyl transferase, HDL cholesterol analytes were stable over 30 days under 22°C, 4°C, and -66°C conditions. This study might be helpful in interpreting the results reflecting the rate of change when inadequate specimens are measured.

Keywords: Chemistry, Stability, Storage, Temperature

Corresponding author: Chang-Eun Park
Department of Biomedical Laboratory Science,
Namseoul University, Cheonan 331-707,
Korea
Tel: 82-41-580-2722
E-mail: pce@nsu.ac.kr

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/3.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

Copyright © 2014 The Korean Society for Clinical Laboratory Science. All rights reserved.

Received: November 27, 2014
Revised: December 28, 2014
Accepted: December 29, 2014

Introduction

The results of biochemistry analytes were changing from the baseline concentration of the fresh sample, according to differentiated temperature and time. The sample treatment of pre-analysis was recommended to conduct in accordance with the guidelines such as an appropriate container, suitable storage, and pre-treatment after blood collection (Ehret *et al*, 2002; CLSI, 2010).

Since the analysis equipment, reagent, the blood collection tube, and sample storage condition differ from laboratory to laboratory, it will be very helpful if we have information about these changing factors. Also, acceptable change limits (ACL) are required for each analytes. In the previous studies, the stability from the baseline concentrations had been conducted under different storage temperature, analysis time, and testing method and analytical ACL had been tried as well (Evans *et al*, 2001; Boyanton *et al*, 2002; Ellis *et al*, 2003).

Each laboratory requires the interpretation reflecting the amount of time change. However, sufficient information is

not provided for specific cases. This study was designed to investigate the changing results according to various time and storage temperature conditions.

Materials and Methods

1. Specimen Preparation and Method

Whole blood was collected in the BD SST II Advance Tube (Becton, Dickinson and Company, Franklin Lakes, NJ, USA). After 30 min for blood clotting for serum samples, the blood was centrifuged at 1500 g for 10 min. About 20 serum samples were fully mixed for pool serum, put 0.5 mL aliquot in tubes, sealed with a cover to prevent evaporation, and stored at 22°C, 4°C, and -66°C temperature conditions. 21 analytes (Table 1) for each pool serum were measured twice on a daily basis up to 30 days using clinical chemistry analyzer Olympus AU2700 (Beckman Coulter, Tokyo, Japan). This study was approved by Ethics Committee of our Institution (Yongin Severance Hospital, Yonsei University, Republic of Korea).

Table 1. Analytes and analytical method

Analytes	Unit	Methods
Calcium	mg/dL	Cresolphthalein complexone
Phosphorus	mg/dL	Phosphomolybdate, UV
Glucose	mg/dL	Hexokinase, UV
BUN	mg/dL	Urease with GLDH (Coupled Enzymes)
Creatinine	mg/dL	Kinetic alkaline picrate (Jaffe reaction)
Uric acid	mg/dL	Uricase
Cholesterol	mg/dL	Enzymatic
Total protein	g/dL	Biuret method
Albumin	g/dL	Dye Binding-BCG
Alkaline phosphatase	IU/L	PNPP, AMP buffer
Total bilirubin	mg/dL	Diazonium salt/Diazonium ion with blank
AST	IU/L	UV without P5P
ALT	IU/L	UV without P5P
Lactate dehydrogenase	IU/L	pyruvate to lactate
γ -glutamyl transferase	IU/L	G-glutamyl-carboxy-nitroanilide (IFCC, 37°C)
Triglycerides	g/dL	Enzymatic without glycerol blank without sample blank
Sodium	mmol/L	ISE, diluted (indirect)
Potassium	mmol/L	ISE, diluted (indirect)
Chloride	mmol/L	ISE, diluted (indirect)
HDL cholesterol	mg/dL	Enzymatic Immunoinhibition
LDL cholesterol	mg/dL	Selective protection method

The measured by chemistry analyzer with Olympus AU2700 (Beckman Coulter, Tokyo, Japan).

2. Statistical Analysis

The changing rate from the initial baseline values was shown in the Levy-Jennings chart. ACL was determined by bi-directional changes based on coefficient of variation (CV), which has the 95% of confidence interval values shown as ISO 5725-6 (ISO, 1994). The CV was obtained from internal quality control values, which had been accumulated for over six months. When the results for an analyte were beyond the limit, they were judged as they had significance, that is, they were the unstable samples.

Results

Stability of individual analyte in pool serum showed various patterns on different periods of time and temperatures shown in Fig. 1. Analytes, of which result values had increased, tended to show far more increased values as specimen were stored under higher temperatures (Fig. 1A). Analytes, whose result values had decreased, tended to be more stable than those at 22°C condition as specimens were stored under lower temperatures. ALP, which showed more or less decreased, were somewhat flexible in temperature

conditions. LDH seemed to be stable at -66°C, but more unstable at 4°C than at 22°C (Fig. 1B).

Most of the 12 analytes such as phosphorus, triglyceride, calcium, BUN, cholesterol, potassium, chloride, creatinine, uric acid, sodium, LDL cholesterol, and total protein increased after 3 to 5 days when stored at 22°C temperature condition. Only 3 analytes such as phosphorus, triglyceride, and calcium were decreased after 23 to 29 days when stored at 4°C condition, however, the rest of the analytes were stable for over 30 days at -66°C condition (Table 2).

Five analytes such as total bilirubin, ALT, AST, alkaline phosphatase, and lactate dehydrogenase were decreased after 1 to 26 days when stored 22°C condition. Pool serum decreased after 1 to 12 days at 4°C condition. Only AST and ALT analytes stable at -66°C condition. Four analytes such as glucose, albumin, γ -glutamyl transferase, and HDL cholesterol were stable during the whole test period at all temperatures (Fig. 1C).

Discussion

In laboratory, samples are rerun after the initial tests are completed, or go through additional tests when necessary.

A

Fig. 1. Stability of routine chemistry analytes according to different time and temperature condition for 30 days. Square, room temperature; triangle, 4°C; circle, -66°C. (A) Increased analytes. (B) Decreased analytes. (C) Stable analytes.

Fig. 1. Continued.

The results might be affected by the storage and time conditions or test reagent. In this study, the stability of the 21 common chemistry analytes shown in Table 1 was evaluated according to different temperature conditions.

When samples were stored at room temperature,

phosphorus, triglyceride, calcium, BUN, cholesterol, potassium, chloride, creatinine, uric acid, sodium, LDL cholesterol and total protein increased at a consistent rate with the lapse of time, and changing range from the baseline values decreased at a low temperature. In particular,

Table 2. Stability of routine chemistry analytes in acceptable changes limit

Analytes	Baseline values	CV% (ACL*)	Out of range after days*		
			22°C	4°C	-66°C
Phosphorus	3.8	1.72 (3.7~3.9)	3 ▲	23 ▲	-
Triglycerides	124	1.53 (120~128)	3 ▲	27 ▲	-
Calcium	8.9	1.59 (8.6~9.2)	4 ▲	29 ▲	-
BUN	12.9	2.48 (12.3~13.5)	9 ▲	-	-
Cholesterol	202	3.11 (189~215)	19 ▲	-	-
Potassium	4.2	1.54 (4.1~4.3)	21 ▲	-	-
Chloride	105	1.4 (102~108)	21 ▲	-	-
Creatinine	0.88	2.16 (0.84~0.92)	22 ▲	-	-
Uric acid	4.9	1.55 (4.7~5.1)	22 ▲	-	-
Sodium	141	1.7 (136~146)	23 ▲	-	-
LDL cholesterol	142	2.69 (134~150)	24 ▲	-	-
Total protein	7.3	1.63 (7.1~7.5)	28 ▲	-	-
Total bilirubin	0.9	2 (0.9~0.9)	1 ▼	1 ▼	1 ▼
ALT	28	2.59 (27~29)	3 ▼	7 ▼	29 ▼
AST	28	2.44 (27~29)	3 ▼	12 ▼	-
Alkaline phosphatase	65	3.14 (61~69)	13 ▼	6 ▼	6 ▼
Latate dehydrogenase	327	3.53 (304~350)	26 ▼	2 ▼	-
Glucose	97	1.58 (94~100)	-	-	-
Albumin	4.5	1.74 (4.3~4.7)	-	-	-
γ -glutamyl transferase	25	2.05 (24~26)	-	-	-
HDL cholesterol	56	3.19 (52~60)	-	-	-

▲ ▼, increase or decrease. *Acceptable change limit was determined by the 95% of confidence interval value for bi-directional changes base on coefficient of variation by ISO 5725-6. (ISO, 1994).

phosphorus, triglyceride and creatinine increased abruptly at room temperature. However, most of the analytes showed stable results in the frozen condition of -66°C for up to over 30 days.

In previous studies, phosphorus, triglyceride, total bilirubin, ALT and AST have been reported unstable at room temperature for 1~3 days (Ehret *et al*, 2002), but calcium, creatinine and uric acid have been reported stable at -20°C up to 4 months (Wilson *et al*, 1972; Donnelly *et al*, 1995). We think that the increased result values in a metabolite group is probably caused by evaporation. Accordingly, a frozen condition or sealing the tubes is recommended for long term storage if possible.

The total bilirubin, ALT, AST, alkaline phosphatase and latate dehydrogenase level decrease consistently with the lapse of sample storage time. The changing rate decreases as the temperature gets lower, but time is considered more influential than the temperature. Alkaline phosphatase was less affected by the temperature conditions. Especially, Latate dehydrogenase is more unstable in the frozen condition than at room temperature, which is more evident in the enzyme

group. In previous studies, total bilirubin, ALT, and alkaline phosphatase has been reported unstable at room temperature (Hwang *et al*, 2002).

Glucose, albumin, γ -glutamyl transferase, HDL and cholesterol were evaluated to be within the acceptable change limits in all the conditions for 30 days. These tests seem stable if they are prevented from being evaporated after being separated as the serum. LDL cholesterol and HDL cholesterol were confirmed more stable for about one month (23~30 days) compared with the data that were resulted within 1~2 days at room temperature in the previous study (Ehret *et al*, 2002). In the previous study, glucose and uric acid analytes decreased under room temperature and refrigerator condition (Ahmed MD *et al*, 2010). However, in our study, glucose did not show any distinguishable changes when serum was separated and uric acid increased, which means we need to consider specimen sealing, collection tubes, storage temperature changes, and specific possible factors in each laboratory.

When the results from the samples in different storage conditions and time are analyzed, they need to be interpreted

according to Fig. 1. The results of this study may be very helpful for the accurate interpretation of clinical reports and determining the sample storage period in laboratories.

Acknowledgements: None

Funding: None

Conflict of interest: None

References

1. Ahmed MD, Daleel AA, Kamal EA. Effect of storage time and temperature on some serum analytes. *International Journal of Pathology*. 2010, 8:68-71.
2. Boyanton BL, Blick KE. Stability studies of twenty-four analytes in human plasma and serum. *Clin Chem*. 2002, 48:2242-7.
3. Clinical and Laboratory Standards Institute (CLSI). Procedures for the handling and processing of blood specimens for common laboratory tests; approved guideline. Document H18-A4, Vol. 30 N°10; 2010.
4. Donnelly JG, Soldin SJ, Nealon DA, Hicks JM. Stability of twenty-five analytes in human serum at 22 degrees C, 4 degrees C, and 220 degrees C. *Pediatr Pathol Lab Med*. 1995, 15:869-74.
5. Ehret W, Heil W, Schmitt Y, Töpfer G, Wissner H, Zawta B, *et al*. Use of anticoagulants in diagnostic laboratory investigations and stability of blood, plasma and serum samples. WHO/DIL/LAB/99.1 Rev 2; 2002.
6. Ellis MJ, Livesey JH, Evans MJ. Hormone stability in human whole blood. *Clin Biochem*. 2003, 36:109-12.
7. Evans MJ, Livesey JH, Ellis MJ, Yandle TG. Effect of anticoagulants and storage temperatures on stability of plasma and serum hormones. *Clin Biochem*. 2001, 34:107-12.
8. Hwang SY, Kim JS, Choo SK, Kim IT, Chang CS. Evaluation of Stability of Quality Control Materials used in Clinical Chemistry Tests. *Korean J Clin Lab Sci*. 2000, 32:177-186
9. ISO 5725-6. Accuracy (trueness and precision) of measurement methods and results. Part 6: use in practice of accuracy values; 1994.
10. Wilson SS, Guillan RA, Hocker EV. Studies of the stability of 18 chemical constituents of human serum. *Clin Chem*. 1972, 18:1498-503.